



ZHONGHONG PULIN MEDICAL PRODUCTS CO.,LTD

Add: West Industrial Park, Tangshan City, Hebei Province, China

Tel: 86-10-67866277

TECHNICAL SPECIFICATION

NITRILE EXAMINATION GLOVES with ZHPFN02

Manufacturer: Zhonghong Pulin Medical Products Co Ltd.

1. Material: nitrile
2. Type: Non-powdered, non-sterile
3. Latex free and finger textured
4. Descripon: double sided, smooth with finger texture
5. Cu□: rolled up
6. Use: disposable
7. Color: blue, violet
8. Packaging: 100 pieces per box
9. High sensivity on the fingers and in the palm, reinforced relief on the tips of the fingers
10. Breaking strength before testing: 6N
11. Suitable for both left and right hands
12. Size: XS-XXL
13. Length of gloves: 240 mm
14. Safe when handling food
15. Protection against the intrusion of microorganisms
16. Resistant to chemicals, acids and non-ketonic solvents
17. Material not causing an allergic reaction
18. They do not contain thiurams and mercaptobenztsols
19. Accelerators used: ZDBC – zinc dubutylidithiocarbamate

Compliance with:

According to Parliament and Council Regulation 2017/745 for Class I medical devices and harmonized

Standards EN 455-1: 2020, EN 455-2: 2015, EN 455-3: 2015, EN 455-4: 2009, EN ISO 10993-1, EN ISO 2859-1

Regulation (UE) 2016/425 for cat. III. personal protective equipment and harmonized standards EN 21420, and EN ISO 374-1: 2016 type B (K, P, T), EN 374-2, EN ISO 16523-1, EN 374-4, EN ISO 374-5: 2016 Bacteria and fungi, EN ISO 374-5: 2016 VIRUS, ASTM D6319-5

In accordance with the EU Food Regulation 1935/2004 part for rubber products: European Councils

RES AP (2004) 4

Marking: CE 2777

Quality management system ISO 9001: 2015 and ISO 13485: 2016

AQL 1.5

Date and place of issue:

Tangshan, 26.02.2023

